

## Technical specification

Weight	120 kg (without vaporizer and cylinder)
Dimensions	1375(H) X 880(W) X 620(D) mm
Display	7 inch color TFT screen
Power supply	100-240V, 50/60Hz
Backup battery	90min (one battery), 150min (two batteries)
Gas supply	O <sub>2</sub> , O <sub>2</sub> +N <sub>2</sub> O, O <sub>2</sub> +Air, O <sub>2</sub> +N <sub>2</sub> O+Air
Backup cylinder yokes	Opt, maximum 2
Flow meter	Dual tubes, Air/O <sub>2</sub> : 0-1L/min, 1-15L/min, N <sub>2</sub> O: 0-1L/min, 1-10L/min
Oxygen flush	25-75 L/min
CO <sub>2</sub> canister	1.5 L
Position for vaporizer	1 or 2
Vaporizer mounting system	Selectatec with interlocking function
Pressure gauge	-20cmH <sub>2</sub> O-100cmH <sub>2</sub> O
Ventilator	Electronic control
Ventilation mode	VCV, PCV, Spontaneous, Manual, SIMV (opt), PSV (opt)
Tidal volume	20-1500 ml (VCV) , 5-1500 ml (PCV)
Monitor parameter	Tidal volume, O <sub>2</sub> concentration, airway pressure, breathing frequency, minute volume, PEEP
Waveform	P-T, F-T, V-T
Alarm	Low airway pressure, High airway pressure, Oxygen failure alarm, Apnea alarm, Breathing rate alarm, O <sub>2</sub> concentration
Oxygen sensor	Opt
CO <sub>2</sub> monitoring	Opt, side-stream
ACGO	Opt
AGSS	Opt, Passive or Active

## WATO EX-20Vet

Anesthesia Machine

# Fully automatic inhalation anesthetic solution for Veterinary





All the essential functions find in our integrated veterinary anesthesia machine

## Design bases on simple facts. We help to make your daily work easier

**EX-20Vet** is designed for the demanding veterinary environment, we know you value animal life as much as human being. so do we dedicate ourselves to design machine that is safe for veterinary Inhalation anesthesia use.

**Automatic**-unlike most of the veterinary anaesthesia machines available in the market which always come with pneumatic control ventilator, EX-20Vet comes with a fully electronic control ventilator that allows users to choose the ventilation modes that best fit for the veterinary needs.

**Simple**-with our highly praised WATO anaesthesia user interface, you will find it extremely easy to use. All the functions and parameters settings arrangement are designed base on clinical opinions and you won't find yourself loss in complex menu. Automatic system leak test help you perform the pre-use check easier and more quickly.

**Precise**-with our highly precised and electronic feedback Pressure control ventilation, you can even ventilate animals that require tidal volume down to 5ml which is ideally for small animals. On top of that EX-20Vet also support ACGO system. The semi close circuit system allows you to take control the whole anaesthesia ventilation process.



# WATO EX-20Vet

## Veterinary Anesthesia Machine



### Physical Specifications

#### Dimensions and Weight

Height	1375 mm
Width	715 mm (without breathing system) 880 mm (with breathing system)
Depth	620 mm
Weight	<120 kg (without vaporizers and cylinders)

#### Top Shelf

Weight limit	30 kg
Length	550 mm
Width	265 mm

#### Work Surface

Height	820 mm
Width	500 mm
Depth	310 mm

#### Drawer (Internal Dimension)

Height	171 mm
Width	390 mm
Depth	315 mm

#### Casters

Diameter	125 mm
Brakes	All four casters with brakes

#### Screen

Display type	Color active matrix TFT
Display size	7.0 inch
Pixel format	800 x 480
Display parameters	All setting and alarm parameters (including Breath rate, I/E ratio, Tidal volume, Minute volume, PEEP, MEAN, PEAK, PLAT, and O <sub>2</sub> concentration)
Display waveforms	P-T, F-T, V-T

### Ventilator Specifications

#### Modes of Ventilation

Manual/Spontaneous Ventilation/Bypass  
Volume Control Ventilation (VCV) with PLV function  
Pressure Control Ventilation (PCV)  
Synchronized Intermittent Mandatory Ventilation (SIMV-Volume Controlled and SIMV-Pressure Controlled)  
Pressure Support Ventilation (PS) with apnea backup

#### Compensation

Automatic circuit leakage and compliance compensation

#### Ventilation Parameters Range

Tidal volume	20~1500 mL ( (Volume Mode) 5~1500 mL (Pressure Mode)
Pinsp	5~60 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O)
Plimit	10~100 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O)
Rate	4~100 bpm (increments of 1 bpm)
I:E	4:1 - 1:8 (increments of 0.5)
Inspiratory pause	OFF, 5% - 60% (increments of 5%)

#### Positive End Expiratory Pressure (PEEP)

Type	Integrated, electronic controlled
Range	OFF, 4~30 cmH <sub>2</sub> O (increments of 1 cm H <sub>2</sub> O)

#### Ventilator Performance

Driving pressure	280 kPa to 600 kPa
Peak gas flow	120 L/min + Fresh Gas Flow

#### Monitoring Parameters

Minute volume	0 ~ 100 L/min
Tidal volume	0~2500 ml
Inspired oxygen (FiO <sub>2</sub> )	18% ~ 100%
Peak airway pressure	-20 ~ 120 cmH <sub>2</sub> O
Mean pressure	-20 ~ 120 cmH <sub>2</sub> O

Plateau pressure	-20 ~ 120 cmH <sub>2</sub> O
PEEP	0 ~ 70 cmH <sub>2</sub> O
Sweep speed	12.5 or 6.25 mm/s

### Control Accuracy

Volume delivery	< 75 ml: ± 15 ml ≥ 75 ml: ± 20 ml or ± 10% of the set value, whichever is greater
Plimit	± 4.0 cmH <sub>2</sub> O or ± 10% of the set value, whichever is greater
Pinsp	± 3.0 cmH <sub>2</sub> O or ± 8% of the set value, whichever is greater
ΔPsupp	± 3.0 cmH <sub>2</sub> O or ± 8% of the set value, whichever is greater
PEEP delivery	± 2.0 cmH <sub>2</sub> O or ± 10% of the displayed value, whichever is greater

### Monitoring Accuracy

Volume monitoring	< 75 ml: ± 15 ml ≥ 75 ml and < 1500 ml: ± 20 ml or ± 10% of the reading, whichever is greater >1500ml: not defined
Pressure monitoring	± 3.0 cmH <sub>2</sub> O or ± 8% of the reading, whichever is greater
PEEP monitoring	0 to 30 cmH <sub>2</sub> O: ± 2.0 cmH <sub>2</sub> O or ± 10% of the reading, whichever is greater >30 cmH <sub>2</sub> O: not defined
MV monitoring	0 to 30L/min: ± 1L/min or ± 15% of the displayed value, whichever is greater Other range: not defined

### Trend Chart

Continuous trend information together with time discrete events are stored and shown by lines for the latest 24 hours with 5 seconds resolution for Tve, Ppeak, MV, Pplat, PEEP, Pmean, Rate and optional FiO<sub>2</sub>. New trend chart will be recorded when restart the machine

### Trend Table

Continuous trend information together with time discrete events are stored and shown by table for the latest 24 hours for TVe, Ppeak, MV, Pplat, PEEP, Pmean, Rate and optional FiO<sub>2</sub>

Resolution 30s, 1min, 2min or 4min optional

New trend form will be recorded when restart the machine

### Alarm Setting

Tidal volume	Low: 0 ~ 1595 ml High: 5 ~ 1600 ml
Minute volume	Low: 0 ~ 99 L/min High: 0.2 ~ 100 L/min
Inspired oxygen	Low: 18% ~ 98% High: 20% ~ 100%
Apnea alarm	20s
Low airway pressure	0 ~ 98 cmH <sub>2</sub> O
High airway pressure	2 ~ 100 cmH <sub>2</sub> O

### Vaporizers

Vaporizer	Mindray V60 Anesthetic Vaporizer or Penlon Sigma Delta Anesthetic Vaporizer
Support agents	Halothane, Enflurane, Isoflurane, Sevoflurane
Position	MAX.2
Mounting mode	Selectatec®, with interlocking function Plug-in®, with interlocking function

### Electrical Specifications

#### Current Leakage

100 ~ 240V	< 500 μA
------------	----------

#### Power And Battery Backup

Power input	without isolation transformer: 100-240 Vac, 50/60 Hz, 6.2~2.6A 100-120 Vac, 50/60 Hz, 5.6A with isolation transformer: 100-120 Vac, 50/60 Hz, 5.6A 220-240 Vac, 50/60 Hz, 2.7A
Battery backup	90 min for 1 piece battery (powered by new fully-charged batteries with 25°C ambient temperature) 150 min for 2 pieces battery (powered by new fully-charged batteries with 25°C ambient temperature)
Battery type	Build-in Li-ion battery, 11.1 VDC, 4400 mAh
Number of batteries	1 or 2 pieces
Time to shutdown	5 min at least (powered by new fully-charged batteries after the first low-power alarm)
Power cord	5 m

#### Auxiliary output supply

Output voltage	220 to 240 V, 100 to 120 V
Output frequency	50/60 Hz
Output current	220 to 240 V : 0.6 A 100 to 120 V : 1.2 A
Fuse	T2AH/250V

## Interface

Wire network	RJ 45 connector 100-Base-TX support upgrading of main unit
--------------	---

## Pneumatic Specifications

### ACGO (Auxiliary Common Gas Outlet)

Connector	ISO 22 mm OD and 15 mm ID
-----------	---------------------------

The outlet locates at the inspiratory limb

### Gas Supply

Pipeline input range	0.28~0.6MPa
Pipeline connections	NIST, DISS
Cylinder input	PISS, Maximum 2 cylinders, optional
Primary regulator nominal output:	207kPa

### O<sub>2</sub> Controls

Method	N <sub>2</sub> O shut off with loss of O <sub>2</sub> pressure
Supply failure alarm	≤ 220.6 kPa
O <sub>2</sub> Flush	25 ~ 75 L/min

### O<sub>2</sub>-N<sub>2</sub>O Link system

Type	Mechanical
Range	Provides a nominal minimum 25% concentration of oxygen in O <sub>2</sub> /N <sub>2</sub> O mixture

### Mechanical Control Flow Meters

O <sub>2</sub> flow range	Two flow tubes with the ranges of 0 ~ 1 L/Min and 1 ~ 15 L/min
Air flow range	Two flow tubes with the ranges of 0 ~ 1 L/Min and 1 ~ 15 L/min
N <sub>2</sub> O flow range	Two flow tubes with the ranges of 0 ~ 1 L/Min and 1 ~ 10 L/min
Accuracy	± 10% of the indicated value (under 20°C and 101.3 kPa, for flow between 10% and 100% of full scale)

### Auxiliary O<sub>2</sub> Flowmeter (optional)

Range	0 ~ 15 L/min
Indicator	Flow tube

### Oxygen Sensor (optional)

Type	Galvanic fuel cell
FiO <sub>2</sub> displayed	18% to 100%
Accuracy	± (volume fraction of 2.5 % +2.5 % gas level)
Response Time	≤20 seconds

## Environmental Specifications

### Operating

Temperature	10 ~ 40°C
Relative humidity	15% ~ 95% (noncondensing)
Barometric (Kpa)	70 ~ 106 kPa

### Storage

Temperature	-20 ~ 60°C for main unit, -20 ~ 50°C for O <sub>2</sub> sensor
Relative humidity	10% ~ 95% (noncondensing)
Barometric	50 ~ 106 kPa

## Breathing Circuit Specification

### Breathing system volume

Automatic ventilation	2600 ml
Manual ventilation	1800 ml
Operational Modes	closed and semi-closed circuit system
Volume of CO <sub>2</sub> canister	about 1500 mL
Water Trap	6 mL, easy to be disassembled

### Breathing Circuit Parameters

Compliance	≤4 mL/100Pa (bag mode) Automatically compensates for compression losses within the breathing circuit in mechanical mode
Expiration resistance	< 6.0 cm H <sub>2</sub> O @60 L/min
Inspiration resistance	< 6.0 cm H <sub>2</sub> O @60 L/min

### System Pressure Gauge

Range:	-20 ~ 100 cmH <sub>2</sub> O
Accuracy:	± (2% of the full scale reading + 4% of the actual reading)

### Ports and Connectors

Exhalation	22 mm OD / 15 mm ID conical
Inhalation	22 mm OD /15 mm ID conical
Manual bag port	22 mm OD /15 mm ID conical

### Bag-to-Ventilator Switch

Type	Bi-stable
Control	Switch between manual and mechanical ventilation

### Adjustable Pressure Limiting (APL) Valve

Range	1 ~ 75 cmH <sub>2</sub> O
-------	---------------------------

Tactile knob indication at above 30 cmH<sub>2</sub>O

Accuracy  $\pm 10$  cmH<sub>2</sub>O or  $\pm 15\%$  of the setting value,  
which is greater

Start pressure  $\leq 2$  cmH<sub>2</sub>O

### Anesthetic Gas Scavenging System (AGSS)

Type of disposal system Active: High-flow or Low-flow

Passive

Size (H x W x D) Active: 430 x 132 x 114 mm

Applicable standard ISO 80601-2-13

Pump rate 75 ~ 105 L/min (High-flow)

25 ~ 50 L/min (Low-flow)

Pressure relief device: Pressure compensation opening to the air

State indication of the disposal system: The float falls below the "MIN" mark on the sight glass when the disposal system does not work or the pump rate is lower than 25 L/min (Low-flow) or 75 L/min (high-flow).

Filter Stainless screen with hole diameter of

140 ~ 150  $\mu$ m

Connector of the disposal system: ISO 9170-2

### Materials

All materials in contact with exhaled patient gases are autoclavable and natural latex free, except flow sensors, O<sub>2</sub> sensor, and mechanical pressure gauge.

-----  
Please contact your local Mindray sales representative for the most current information.

[www.mindray.com](http://www.mindray.com)

P/N:ENG-WATO EX-20 Vet Datasheet -210285X4P-201904020

©2019 Shenzhen Mindray Bio-Medical Electronics Co.,Ltd. All rights reserved.

**mindray**  
healthcare within reach